

Stroke and **Medical Devices**

Carlos Peña, PhD, MS

Director

Division of Neurological and Physical Medicine Devices

Food and Drug Administration

October 6, 2015

Treatments

Pharmacotherapy Medical Devices

Regulatory **Decision Points**

Medical Device Classification


Benefit Risk Assessment

Predicate Devices

Valid Scientific Evidence

Least Burdensome Approach

Current State of Review






When is Clinical Data Needed?


- PMA: almost always needed
- De novo: typically needed, but not always
- 510(k): generally not needed

You can request feedback on any protocols through the pre-submission process, preferably before starting the study


Experience in Moving Neurological Medical Devices From **Bench to Market**



Cleared Mechanical Neurothrombectomy (Clot Retrieval) Devices




MERCI Retriever



Penumbra Reperfusion System



Medtronic/Covidien
Solitaire
Revascularization Device




Stryker Neurovascular
Trevo ProVue Retriever

Intended Use: Restoring blood flow or revascularization in intracranial vessels within 8 hours of acute ischemic stroke symptom onset. For patients who are ineligible to receive or who failed intravenous tissue plasminogen activator (IV t-PA).

Center for Devices and Radiological Health (CDRH) Organization

Pathway for Neurological and Physical Medicine Regulatory Submissions



Neurothrombectomy Guidance
Guidance for Industry and FDA Staff
Pre-Clinical and Clinical Studies
for Neurothrombectomy Devices
Document issued on: June 18, 2007

Guidance Topic Areas

Biocompatibility

Pre-Clinical Evaluation

Bench Testing

Animal Testing

Clinical Studies

Measures of Success

Alternate Sources of Clinical Info

Meeting Goals-Devices

4 Topic Areas

Clinical Study Design, Patient Groups

Clinical Outcomes, Imaging, Safety

Statistical Considerations

Registries

Closing Thoughts